



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
2098 Gaither Road
Rockville MD 20850

MAY 20 2004

HemoCue, Inc.
c/o Mr. R. J. Slomoff
Consultant to HemoCue, Inc.
9229 Cranford Drive
Potomac, MD 20854

Re: k041230
Trade/Device Name: HemoCue® Glucose 201 DM System
Regulation Number: 21 CFR 862.1345
Regulation Name: Glucose test sytem
Regulatory Class: Class II
Product Code: LFR
Dated: May 6, 2004
Received: May 10, 2004

Dear Mr. Slomoff:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820).

Page 2

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific information about the application of labeling requirements to your device, or questions on the promotion and advertising of your device, please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 594-3084. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>.

Sincerely yours,



Jean M. Cooper, MS, D.V.M.

Director
Division of Chemistry and Toxicology
Office of *In Vitro* Diagnostic Device
Evaluation and Safety
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): **K041230**

Device Name: **HemoCue® Glucose 201 DM System**

Indications For Use:

Quantitative determination of glucose in whole blood using a specially designed analyzer, the HemoCue Glucose 201DM with Plasma conversion. The quantitative determination of the instant blood glucose concentration in circulation supplements the clinical evidence in the diagnosis and treatment of patients with diabetes as well as monitoring of neonatal blood glucose levels.

Thereagents/microcuvettes and the analyzer form an analytical system. The HemoCue Glucose 201DM Analyzer is only to be used with HemoCue Glucose 201 Microcuvettes. The use of any other device in the HemoCue Glucose 201DM Analyzer except the HemoCue Glucose 201 Microcuvettes is neither supported nor recommended by HemoCue, and could give erroneous results with serious clinical consequences.

The HemoCue Glucose 201 Microcuvettes are for in vitro use only.


Prescription Use XXX
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)


Division Sign-Off

Office of In Vitro Diagnostic Device
Evaluation and Safety

510(k) K 041230

Page 1 of 1